Analytical Reference Materials International

Certificate of Analysis Certified Reference Material

Grade: Alloy 255 / UNS S32550
Part Number (Q.A. NO.): IARM 239A

Certificate Date: 04/18/2002 Certificate No.: 239A-04182002-IARM-F Revision Date: 05/19/2003

Interpretation of Data

- 1. Certified values listed below reflect analysis results submitted by qualified analytical laboratories using a combination of methods and instrumentation that emulate actual methods and instrumental techniques currently utilized in the analytical community.
- 2. Any data reported and enclosed by a **parentheses ()** is a "best estimate" and is NOT CERTIFIED. This data could not be quantified sufficiently for certification. It was however, reported by enough laboratories to be considered as potentially present in the matrix of the material being examined.
- 3. The "Inter-laboratory Analysis Program" (ILAP) utilized in the establishment of the data are an ongoing program with permanent membership. Certain elements may be selected by a consensus of the members for more extensive testing. Therefore the data in **brackets** [] **indicates further testing is in process.**
- 4. The "± Confidence Interval at 95%" is enclosed by a parentheses () below the individual element's concentration.

IMPORTANT: A "User Registration Card" accompanies all shipments. This card should be completed immediately upon receipt of materials with the appropriate user information. This is the only way in which ARMI can guarantee customer updates or possible data modifications!

Aluminum 0.005 (0.003)	Boron 0.0030 (0.0004)	Carbon 0.028 (0.002)	Cobalt 0.037 (0.002)	Chromium 25.74 (0.08)	Copper 1.63 (0.03)	Manganese 1.01 (0.02)
Molybdenum 3.14 (0.03)	Nitrogen 0.253 (0.005)	Niobium 0.010 (0.002)	Nickel 6.07 (0.04)	Oxygen 0.017 (0.002)	Phosphorus 0.023 (0.001)	<u>Sulfur</u> 0.0015 (0.0003)
Silicon 0.40 (0.01)	<u>Tin</u> <0.01	<u>Titanium</u> 0.003 (0.0004)	Vanadium 0.05 (0.004)	Tungsten 0.027 (0.004)		

The laboratories participating in the "Inter-Laboratory Analysis Program" (ILAP) and certification of this material are as follows:

AK Steel, Butler Works - Butler, PA
Allegheny Ludlum Corp. - Brackenridge, PA
Anderson Laboratories, Inc. - Greendale, WI
Colorado Metallurgical Services - Denver, CO
Jorgensen Forge Corp. - Seattle, WA
MetalTek International, Inc. - Waukesha, WI

Allvac Lockport - Lockport, NY
Carpenter Technology Corporation - Reading, PA
Conam Inspection - Gary, IN
Laboratory Testing, Inc. - Hatfield, PA
Timken Latrobe Steel Co. - Latrobe, PA

Allegheny Ludlum Corp. - Brackenridge, PA

Traceability: All members of the "Inter-Laboratory Analysis Program" (ILAP) listed above validate test methods and instrument performance utilizing SRMs produced by the National Institute of Standards and Technology, (NIST) as well as other CRMs and RMs produced by recognized Certifying Bodies from around the world. The specific SRMs, CRMs and RMs applicable to the material covered by this certificate are: NIST 3101A, 3107, 121D, 3113, 1336, 3137, 133B, 3149, 3155, 3162A, 3163, 3103A, 3106, 3109A, 3131A, 3128, 3156, 3169, 3102A, LECO 501-643, IARM 212A, NIST 362, IS 0077A, SS 462, 466/1, 467/1, LECO 502-072, 501-647, IARM 27B, 27C, NIST 136E, 897, 898, BS 179A, NIST 368, 1761, 1762, 1763, 1764, 1765, 1766, 1767, 1768, 1261a, 1262b, 1263a, 1264a, 1265a, BAS 401/1, 402/1, 403/1, 404/1, 405/1, 406/1, 407/1, 408/1, 409/1, 410/1, BCS 468, BS 347, LECO 501-550, 501-553, SKS 11, BS 179AWCI H0425, LECO 501-501, 502-016, AR 663, BS 179A, NIST 2159, BS 179, LECO 502-016, NIST 339, 121D, 442, C2401, BS 179A, IARM 2C, 10A, 10B, 15B, 160A, 162A, LECO 501-551, AR 663, BAS 475, BS 179, CA316, LECO 502-016, 502-072

A specific line of traceability is established to NIST and other Certifying Bodies for those elements that are noted as "Certified Values" on the Certificates of Analyses referenced above.

See Reverse Side for Statistical Data and Additional Information Regarding this Material.

The following data and accompanying statements represent all pertinent information reported in the ILAP as it applies to the chemical characterization of this material as of 05/19/2003.

239A	Al	В	С	Со	Cr	Cu	Mn	Mo	N	Nb	Ni	0	P	S	Se
1	0.0002	0.0038	0.023	0.0358	25.85	1.56	1.06	3.163	0.266	0.006	6.086	0.0159	0.023	0.0015	0.001
2	0.010	0.0029	0.028	0.042	25.888	1.680	0.995	3.158	0.255	0.0078	6.044	0.0191	0.022	0.0010	< 0.00001
3	0.0049	0.0037	0.0280	0.035	25.670	1.71	1.015	3.152	0.2526	0.0073	6.059	0.0145	0.0225	0.0021	0.0053
4	0.0055	0.0025	0.0274	0.034	25.563	1.595	1.040	3.207	0.2675	0.0095	5.947	0.0159	0.0239	0.0015	< 0.01
5	0.00751	0.00274	0.0325	0.0327	25.869	1.637	0.9728	3.086	0.2412	0.0090	6.133	0.0190	0.0217	0.0010	
6	0.009	0.0030	0.0235	0.038	25.89	1.630	1.028	3.093	0.2556	0.012	6.16	0.0164	0.0259	0.0013	
7	0.004	0.0028	0.030	0.039	25.590	1.638	1.043	3.200	0.248	0.0144	5.980		0.021	0.0011	
8	0.0003	0.0028	0.0291	0.0360	25.495	1.590	0.990	3.07	0.250	0.014	6.10		0.0222	0.0017	
9		0.0020	0.031		25.85	1.61	0.986	3.19	0.257	0.015	6.08		0.025	0.0013	
10		0.0036	0.0280		25.83	1.65	0.994	3.08	0.2525	0.010	6.09		0.023	0.0021	
11			0.028		25.596	1.654	1.01	3.13	0.2420	0.0082	6.10		0.0255		
12			0.0239		25.858		1.04	3.159		0.0112	6.11		0.0251		
13					25.76		0.980				5.97				
14					25.70										
Mean	0.0052	0.0030	0.0277	0.0366	25.7435	1.6322	1.0118	3.1407	0.2534	0.0104	6.0661	0.0168	0.0234	0.0015	0.0032
STDV.	0.0037	0.0006	0.0030	0.0030	0.1383	0.0425	0.0280	0.0483	0.0084	0.0030	0.0646	0.0019	0.0016	0.0004	0.0030
Certified	0.005	0.0030	0.028	0.037	25.74	1.63	1.01	3.14	0.253	0.010	6.07	0.017	0.023	0.0015	
95% C.I.	0.003	0.0004	0.002	0.002	0.08	0.03	0.02	0.03	0.005	0.002	0.04	0.002	0.001	0.0003	
Methods	D,I,O	D,I,O	C,O	X,I,O	X,W,O	X,I,O	X,I,O	X,I,O	F,O	X,I,O	X,W,O	F	X,I,O	C,O	

Methods: W = Classical, C = Combustion, F = Fusion, A = AA or GFAA, I = ICP or DCP, D = DC Arc, O = OE, X = XRF, G=GDMS

239A	Si	Sn	Ta	Ti	V	W	Ag	As	Bi	Ca	Mg	Pb	Se	Te	Zr
1	0.416	0.004	0.0013	0.002	0.042	0.0205	0.0002	0.007	0.00006	0.0181	0.0001	0.00004	0.00012	0.00008	0.0003
2	0.411	0.0061	0.005	0.0027	0.0543	0.0220		0.00541	0.0025		< 0.0001	< 0.0001			0.0021
3	0.388	0.0025	0.0031	0.0024	0.051	0.030		0.0064	< 0.001		< 0.001	0.00005			0.0008
4	0.383	0.0077	0.001	0.002	0.0534	0.020		0.001				0.001			0.0021
5	0.413	0.00397	0.0022	0.00275	0.065	0.0244		0.013				0.0063			0.002
6	0.385	0.0068	0.008	0.0024	0.0565	0.033		0.0031				0.0003			0.0026
7	0.386	0.0038	0.0229	0.002	0.040	0.034						< 0.001			0.003
8	0.415	0.0028		0.003	0.0511	0.031						0.0026			
9	0.418	0.0077		0.0034	0.052	0.0277									
10	0.404			0.0038	0.057										
11	0.397				0.052										
12	0.3988				0.066										
13					0.0542										
Mean	0.4012	0.0050	0.0062	0.0026	0.0534	0.0270	0.0002	0.0060	0.0013	0.0181	0.0001	0.0017	0.0001	0.0001	0.0018
STDV.	0.0133	0.0020	0.0077	0.0006	0.0073	0.0054		0.0041	0.0017			0.0024			0.0010
Certified	0.40	< 0.01	(0.003)	0.003	0.05	0.027		(0.006)				< 0.005			(0.002)
95% C.I.	0.01			0.0004	0.004	0.004									i
Methods	X,W,I,O	X,A,I,O	D,I,O	X,I,O	X,I,O	X,I,O		D,A,O				X,D,A,O			X,D,I,O

Methods: W = Classical, C = Combustion, F = Fusion, A = AA or GFAA, I = ICP or DCP, D = DC Arc, O = OE, X = XRF, G=GDMS

The International Standards Organization (ISO) definitions, expressed in ISO Guide 30-1981-(E) list the following:

<u>Certifying Body:</u> A technically competent body (organization or firm, public or private) that issues a Reference Material Certificate. The only generally accepted certifying body in the United States is the U. S. Department of Commerce, National Institute of Standards & Technology, (NIST), Gaithersburg, MD.

Reference Material (RM): A material or substance with one or more properties which are sufficiently well established to be used for calibration of an apparatus, the assessment of a measurement method, or for assigning values to materials.

<u>Certified Reference Material (CRM):</u> A reference material with one or more properties whose values are certified by a technically valid procedure accompanied by or traceable to a certificate or other documentation, which is issued by a Certifying Body.

Inter-Laboratory Analysis Program (ILAP): Although ASTM Standard E691-87 applies to inter-laboratory studies to "Determine the Precision of a Single Test Method", it is also a well thought out and logical plan for conducting an inter-laboratory program involving multiple techniques. Therefore, the planning, conducting, analyzing, protocol and treatment of data resulting from this inter-laboratory program were performed utilizing the guidelines established in ASTM E691-87.

<u>Methods of Analysis:</u> In view of the fact, that the "Inter-Laboratory Analysis Program" entails a wide variety of materials, no single analytical method would provide optimum data results. Therefore, the methods utilized were a combination of ASTM Standard Methods for classical wet chemistry, ICP, AA, Optical Emission and X-Ray spectrometric methods. The determinations for Carbon, Sulfur, Nitrogen and Oxygen are the result of combustion instrument procedures.

Selection of Materials: A "batch" or "series" is defined as a single bar of one continuous length. The majority of materials are in wrought condition. Other methods of manufacture are utilized as a last resort, only in the case of those materials being unavailable in wrought condition. "Batch" samples are taken by removing a one-inch cross section for every thirteen inches of total length from the entire bar. Twenty-five percent of the one inch cross section is converted to chips for analysis by classical wet chemistry, ICP, AA, and combustion procedures and seventy-five percent remains in a solid disk form for OES and X-Ray analysis where applicable. Each member of the ILAP is furnished both a solid sample and the corresponding supply of chips from a specific location on the batch bar. This massive sampling procedure results in the homogeneity being reflected as a product of the overall statistics and certified data.

Certified by:

William D. Britt, President & General Manager Analytical Reference Materials International Certificate No.: 239A-04182002-ARM-F Certificate Date: 04/18/2002 Revision Date/No.: 05/19/2003