



Certificate of Analysis

IARM Fe304L-18

Stainless Steel AISI 304L / UNS S30403

Certified Reference Material

Certified Values listed in wt.% with associated uncertainties

Al	0.003 ± 0.001	As	0.007 ± 0.001	B	0.0012 ± 0.0009	C	0.024 ± 0.002
Co	0.156 ± 0.003	Cr	18.34 ± 0.09	Cu	0.54 ± 0.01	Mn	1.39 ± 0.02
Mo	0.462 ± 0.005	N	0.081 ± 0.003	Nb	0.013 ± 0.005	Ni	8.17 ± 0.06
O	0.006 ± 0.002	P	0.034 ± 0.002	S	0.029 ± 0.004	Si	0.43 ± 0.01
Sn	0.013 ± 0.003	Ti	0.0056 ± 0.0009	V	0.076 ± 0.004	W	0.056 ± 0.005

Indicative Values listed in ppm

Ca (20)	Ce (<20)	H (8)	Mg (<10)	Pb (<10)	Sb (20)	Se (<10)
Ta (<50)	Zn (<50)	Zr (<50)				

Description and Intended Use

This CRM may come in the form of a solid disk, chips, or powder. The intended use of this CRM may include, but is not limited to, the calibration of instruments and the validation of analytical methods.

Instructions for Use

1. The test surface is on the opposite side of the labeled surface, which includes the material identification. The entire thickness of the unit is certified. However, the user is cautioned not to measure disks less than 2 mm thick when using X-ray fluorescence spectrometry. Each packaged disk has been prepared by finishing the test surface using a lathe. The user must determine the correct surface preparation procedure for each analytical technique. The user is cautioned to use care when either resurfacing the disk or performing additional polishing, as these processes may contaminate the surface.
2. The minimum sample size for chips should be individually evaluated based on the analytical technique used; this would typically be greater than 0.1 grams.
3. The material should be stored in a cool, dry location when not in use.
4. Chips are not certified for Oxygen analysis.



The following data represents all pertinent information reported as it applies to the chemical characterization of this material.

	Al	As	B	C	Ca	Ce	Co	Cr	Cu	H	Mg	Mn	Mo	N	Nb	Ni
1	0.001	0.0049	0.000456	0.0179	0.0014	<0.002	0.148	18.1133	0.5194	0.000849	0.000310	1.343	0.444	0.0756	0.0005	8.0095
2	0.001	0.00609	0.0006	0.022	0.00179		0.1486	18.195	0.52	0.00038	<0.0001	1.346	0.4549	0.079033	0.0024	8.10
3	0.00129	0.0062	0.0008	0.023	0.0023		0.151	18.20	0.5267	0.00113	<0.0005	1.36	0.455	0.0805	0.0031	8.10635
4	0.002	0.0067	0.000833	0.0232	<0.0005		0.1526	18.2123	0.52767		<0.001	1.374	0.456	0.0807	0.014	8.119
5	0.0024	0.0075	0.0023	0.0245	<0.001		0.156	18.2176	0.534			1.38	0.4601	0.0808	0.016	8.124
6	0.003	0.0086	0.0025	0.0247	<0.005		0.1562	18.27	0.5424			1.38	0.465	0.0816	0.016	8.12667
7	0.0032			0.0249			0.157	18.3133	0.5445			1.40	0.46567	0.0821	0.0168	8.12667
8	0.0042			0.0249			0.157	18.32	0.545			1.402	0.46663	0.088	0.0173	8.137
9	0.0043			0.025			0.157	18.36	0.548			1.403	0.467		0.0181	8.147
10	0.005			0.026			0.15767	18.442	0.55			1.408	0.46733		0.01878	8.2038
11				0.028567			0.158	18.445	0.56301			1.41228	0.468667		0.019	8.232
12							0.16091	18.48	0.564			1.4231	0.471			8.26
13							0.16408	18.55	0.565			1.424	0.471			8.341
14							0.165	18.614	0.569			1.4373				8.353
15																
Mean	0.003	0.007	0.0012	0.024	0.002		0.156	18.34	0.54	0.0008		1.39	0.462	0.081	0.013	8.17
STDV.	0.001	0.001	0.0009	0.003	0.0005		0.005	0.1	0.02	0.0004		0.03	0.008	0.003	0.007	0.1
Certified	0.003	0.007	0.0012	0.024	(0.002)	(<0.002)	0.156	18.34	0.54	(0.0008)	(<0.001)	1.39	0.462	0.081	0.013	8.17
U _{CRM}	0.001	0.001	0.0009	0.002			0.003	0.09	0.01			0.02	0.005	0.003	0.005	0.06
Methods	I,IM,G,O	IM,A,I,O	I,IM,O	C,G,O	I,IM	I	I,O,X,IM,G	I,O,X,G	I,O,X,IM,G	F	IM,I	I,O,X,G	I,O,X,G	C,F	I,IM,G,X,O	I,O,X,G

	O	P	Pb	S	Sb	Se	Si	Sn	Ta	Ti	V	W	Zn	Zr
1	0.0036	0.027	0.0001	0.0186	0.001567	<0.0001	0.397	0.003	0.003	0.0035	0.064	0.034	0.000311	0.0001433
2	0.0045	0.02776	<0.001	0.0264	0.00176	<0.001	0.404	0.0096	<0.001	0.0043	0.06888	0.0469	0.001533	0.000146
3	0.0055	0.0292	<0.001	0.0277	0.0021	<0.0010	0.4167	0.01	<0.0010	0.005	0.0725	0.0512667	<0.0005	0.002
4	0.0055	0.03	<0.0010	0.028	0.003		0.421	0.0144667	<0.002	0.005	0.073	0.0546	<0.001	0.004
5	0.006	0.0321		0.0286	<0.001		0.427	0.0149	<0.005	0.005	0.0747667	0.0547667	<0.0010	<0.0010
6	0.00685	0.03285		0.0292			0.4272	0.015		0.0053	0.07505	0.054833	<0.005	<0.005
7	0.006867	0.033433		0.0292			0.4298	0.015033		0.0058	0.0753	0.0575		
8	0.011	0.033733		0.0298			0.43	0.01585		0.006	0.076	0.059		
9		0.035667		0.03			0.434	0.0169		0.00633	0.076033	0.05945		
10		0.0367		0.0301667			0.441	0.0173		0.0067	0.0763	0.061		
11		0.037		0.0303			0.441667			0.0086	0.0764	0.065		
12		0.038		0.0382			0.4442				0.077	0.06548		
13		0.0389					0.453				0.081	0.068		
14		0.0396					0.454				0.0915			
15														
Mean	0.006	0.034		0.029	0.002		0.43	0.013		0.0056	0.076	0.056	0.001	0.002
STDV.	0.002	0.004		0.004	0.0006		0.02	0.004		0.001	0.006	0.009	0.0009	0.002
Certified	0.006	0.034	(<0.001)	0.029	(0.002)	(<0.001)	0.43	0.013	(<0.005)	0.0056	0.076	0.056	(<0.005)	(<0.005)
U _{CRM}	0.002	0.002		0.004			0.01	0.003		0.0009	0.004	0.005		
Methods	C,F	I,O,X,IM,G	I,IM,A	C,G,X,O	IM,A,I	IM,A,I	I,O,X,G	I,O,X,IM,A	I,IM,G	I,O,X,IM,G	I,O,X,IM,G	I,O,X,IM,G	IM,I	IM,I

Legend: W = Classical, C = Combustion, F = Fusion, A = AA or GFAA, I = ICP or DCP, IM=ICP-MS, D = DC Arc, O = AES, X = XRF, G = GDAES or GDMS, H = Hollow Cathode AES

Certification Laboratories

Connecticut Metallurgical Inc
IMR Test Labs
EAG Laboratories
Luvak Laboratories Inc.
New Hampshire Materials Lab

East Hartford, CT
Lansing, NY
Liverpool, NY
Boylston, MA
Somersworth, NH

Dirats Laboratories
VHG Labs
Laboratory Testing, Inc.
Massachusetts Materials Research Inc.

Westfield, MA
Manchester, NH
Hatfield, PA
West Boylston, MA

Certification laboratories have demonstrated performance and traceability by utilizing test methods under the scope of ISO 17025 or have shown competence through a proficiency testing program. Some of the specific SRMs used in the analysis of the material covered by this certificate are:

NIST 100C

NIST 101G

NIST 152B

NIST 157B

NIST 160B

NIST 363

Homogeneity and Uncertainty

"Uncertainty" values, as reported adjacent to certified concentration values, are based on a 95% Confidence Interval. These estimated uncertainties include the combined effects of method imprecision, material inhomogeneity, and any bias between methods. Homogeneity data from experimental XRF results are reflected in both the overall statistics and certified data. Homogeneity samples are selected by a systematic sampling procedure. The number of samples may be determined by equation 1, where N_{prod} is the number of units produced and N_{min} is the number of samples used for homogeneity testing. These samples are arranged in a simple randomized design such that each sample is analyzed multiple times by XRF. Homogeneity may also be determined within sample using an applied version of ASTM E826. A single factor ANOVA is used to calculate uncertainty due to inhomogeneity (U_{hom}). Uncertainty of the material is calculated by equation 2, where $H=U_{hom}$, S = Standard deviation, t = t-value at 95% CI, and n = number of observations.

$$1. N_{min} = \max(10, \sqrt[3]{N_{prod}})$$

$$2. U_{CRM} = \frac{\sqrt{H^2 + S^2}}{\sqrt{n}} * t$$

Expiration

The certification of this material is valid indefinitely, within the uncertainty specified, provided the material is handled and stored in accordance with the instructions stated on this certificate. The certification is nullified if the material is damaged, contaminated, otherwise modified, or used in a manner for which it was not intended.



David Coler, General Manager

Analytical Reference Materials International



Analytical Reference Materials International • 276 Abby Road • Manchester, NH 03103
Telephone (603) 935-4100 • Fax (603) 935-4101 • www.ARMI.com • ARMI@LGCgroup.com

IARM-Fe304L-02262018

2/28/2018

2 / 2